



Click'aV® and Click'aV Plus™ Ligating Clips
Instructions for use

Ref. no.: 0301-03M, 0301-03ML, 0301-03L, 0301-03XL, 0301-03M04, 0301-03ML04, 0301-03L04, 0301-03XL04, 0301-03ML02, 0301-03L02, 0301-03XL02, 0301-10ML, 0301-10L, 0301-10XL, 0301-10L04, 0301-10ML04, 0301-10XL04, 0301-10L02, 0301-10ML02, 0301-10XL02, 0301-10XXL02, 0301-10XXL03, 0301-10XXL04

 <p>Grena Ltd, 1000 Great West Road, Brentford, Middlesex TW8 9HH, United Kingdom</p>	<p>Contact information: Phone /Fax: + 44 115 9704 800</p>	<div style="display: flex; align-items: center;"> <div style="border: 1px solid black; padding: 2px; margin-right: 5px;">EC</div> <div style="border: 1px solid black; padding: 2px; margin-right: 5px;">REP</div> <div style="margin-left: 20px;"> <p>MDML International Unit 7, Argus House Greenmount Office Park, Harold's Cross Road Dublin 6W, DUBLIN Ireland D6W PP38</p> </div> </div>		<p align="center">ENG IFU-044-ENG_16</p>
--	--	---	---	---

Important: This instruction cannot be used as a manual for surgical techniques used during the work with ligating clips. To learn adequate knowledge about surgical technique it is necessary to contact our company or authorized distributor and to acquaint with appropriate technical instructions, professional medical literature and graduate proper training under supervision of surgeon experienced in techniques of minimally invasive surgery. Before use we recommend reading precisely all information included in this manual. Not being obedient to this information may lead to serious surgical consequences such as patient injury, contamination, infection, cross-infectionability of ligation or death.

Indications: Click'aV® and Click'aV Plus™ ligating clips are intended for ligating of any linear tissue structures or vessels during an operation for haemostasis where use of non-absorbable clips is required. Conformity of the size of the occluded tissue and the clips is required.
Patient target group - adult and young patients, males and females.
Intended users: product is intended to be used exclusively by qualified medical staff.

Contraindications:
DO NOT use for tubal ligation as contraceptive method.
DO NOT use for renal artery ligation during laparoscopic live donor nephrectomy.
DO NOT use as a tissue marker.

Description of the device: Click'aV® and Click'aV Plus™ Ligating Clips are sterile, disposable, non-active, and biocompatible implantable devices developed for use in surgical procedures requiring reliable vessel or tissue ligation. The clips are crafted from a durable, non-absorbable polymer suitable for permanent implantation. Each clip is equipped with a secure latch mechanism, providing dependable closure and enhanced lateral stability. The Click'aV Plus™ model includes an advanced design feature: staggered, inward-facing teeth along the inner clip surface that are angled toward the tissue. This specialized tooth arrangement significantly enhances both lateral and longitudinal grip stability, achieving much higher stability than standard polymer clips, thereby optimizing secure tissue management in challenging surgical environments. Click'aV Appliers are compatible devices used to apply these clips. Designed to facilitate controlled clip application, the appliers allow for precise closure around targeted tissue. Click'aV® and Click'aV Plus™ Ligating Clips are intended for healthcare professionals trained in vessel and tissue ligation techniques. Click'aV® and Click'aV Plus™ Ligating Clips can be opened with specially designed clip remover.

MR Safety Information: Click'aV® and Click'aV Plus™ Ligating Clips are MR Safe. Click'aV® and Click'aV Plus™ Ligating Clips are made of a non-absorbable polymer material and are considered MR Safe. It means they poses no known hazards in all MR imaging environments as they are non-conductive, non-metallic and non-magnetic items.

- Instructions for use:**
- Choose the appropriate size of the clip and the compatible applier.
 - Check the compatibility of all devices prior to use.
 - Following aseptic rules remove clips cartridge from the single packing. To prevent any damage of the device place it on a sterile surface.
 - Grip the applier around the bolt (like pencil is gripped). For endo appliers grip applier around the shaft. This grip ensures that the jaws of the device remain fully open, what is essential for proper clip loading.
 - Align applier jaws vertically and laterally over a clip in the cartridge and advance product jaws into the slot of the clip cartridge making sure they are perpendicular to the surface of the cartridge. Incorrect position of the jaws during loading may lead to incorrect seating of the clip in the jaws, which may result in the inability to securely close the clip, its cracking, deformation or falling out of the applier. Advance the jaws till there is an audible click. Do not use force to push the applier. The applier should move inside and outside of the slot easily. Using excessive force to push the applier may break the clip.
 - Remove the applier from the cartridge. It may be necessary to hold the cartridge to allow the clip to be removed. Make sure the clip is affixed in the jaws securely. The clip bosses should seat in the notches of the applier jaws. Incorrect seating of the clip in the jaws, may result in the inability to securely close the clip, its cracking, deformation or falling out of the applier.
 - Sufficiently skeletonize the structure to be ligated to allow locking mechanism of the clip to be clear of tissue to avoid penetration of the latch through the tissue. Penetration of the tissue by the latch affects closure security, may deform or even break the clip.
 - If endoscopic applier is used squeeze the applier handles (but take care not to lock the clip shut) and insert the applier jaws and shaft down the cannula. Maintain compression on the applier handles until the jaws clear the cannula. This procedure is necessary because cannula inner diameter is in most of the cases smaller than external dimension of opened applier jaws. Squeezing of the applier handles may be also necessary during applier withdrawal from the cannula. If the handle is not sufficiently compressed, the jaws of the applier can scrape the material from the inside of the cannula and detached plastic particles can fall into the body cavities.
 - During application, rotate endoapplier shaft this way that the single tooth of the clip will go down and can be seen from the top and side at a time. This allows the user to visually confirm encapsulation of the structure being ligated. For open surgery it is also recommended that single tooth will be in down position.
 - Position the clip around the structure intended for ligation in a manner that provides clear visualization of the locking mechanism. Use appropriate force to close the clip completely until it locks shut, making sure it is placed properly. Releasing the pressure on the handles will cause the applier jaws to spring open.
 - Remove the applier from the surgical site.


















Click'aV® and Click'aV Plus™ clip size	Compatible Click'aV® clip appliers	Ligated structure size in [mm]
M	0301-04M20, 0301-04M27, 0301-04M27A70, 0301-04ME, 0301-04MEA20, 0301-04MEA45, 0301-04MEA20B, 0301-04MEB, 0301-04MMEHS, 0301-04MMEHSB, 0301-04MEI, 0301-04MEA20I, 0301-04MEA45I, 0301-04MEIB, 0301-04MEA20IB, 0301-04MEA45IB, 0301-04MEHS, 0301-04MEOMN, 0301-04MEOMNB, 0301-04MEOMNBX, 0301-04MEOMNBX, 0301-04MEN, 0301-04MEA20N, 0301-04MEA45N, 0301-04MEBN, 0301-04MEA20BN, 0301-04MEA45BN	2 to 7
ML	0301-04ML20, 0301-04ML27, 0301-04ML27A70, 0301-04MLE, 0301-MLEA20, 0301-MLEA45, 0301-04MLEB, 0301-04MLEA20B, 0301-04MMEHS, 0301-04MMEHSB, 0301-04MLEI, 0301-04MLEA20I, 0301-04MLEA45I, 0301-04MLEIB, 0301-04MLEA20IB, 0301-04MLEA45IB, 0301-04MLEHS, 0301-04MLEOMN, 0301-04MLEOMNB, 0301-04MLEOMNBX, 0301-04MLEOMNBX, 0301-04MLEOMNX, 0301-04MLE, 0301-04MLEN, 0301-04MLEA20N, 0301-04MLEA45N, 0301-04MLEBN, 0301-04MLEA20BN, 0301-04MLEA45BN	3 to 10
L	0301-04L20, 0301-04L27, 0301-04L27A70, 0301-04LE, 0301-04LEA20, 0301-04LEA45, 0301-04LEB, 0301-04LEA20B, 0301-04LXLEHS, 0301-04LXLEHSB, 0301-04LEI, 0301-04LEA20I, 0301-04LEA45I, 0301-04LEIB, 0301-04LEA20IB, 0301-04LEA45IB, 0301-04LEOMN, 0301-04LEOMNB, 0301-04LEOMNBX, 0301-04LEOMNBX, 0301-04LEOMNX, 0301-04LXLE, 0301-04LXLEA20, 0301-04LXLEA45, 0301-04LXLEIB, 0301-04LXLEA20IB, 0301-04LXLEA45IB, 0301-04LXLEHS, 0301-04LXLEOMN, 0301-04LXLEOMNB, 0301-04LXLEOMNBX, 0301-04LXLEOMNBX, 0301-04LXLEOMNX, 0301-04LXLUNE	5 to 13
XL	0301-04XL20, 0301-04XL27, 0301-04XL27A70, 0301-04XLE, 0301-04XLEA20, 0301-04XLEA45, 0301-04XLEB, 0301-04XLEA20B, 0301-04XLEHS, 0301-04XLEHSB, 0301-04XLEI, 0301-04XLEA20I, 0301-04XLEA45I, 0301-04XLEIB, 0301-04XLEA20IB, 0301-04XLEA45IB, 0301-04XLEHS, 0301-04XLEOMN, 0301-04XLEOMNB, 0301-04XLEOMNBX, 0301-04XLEOMNBX, 0301-04XLEOMNX, 0301-04XLXUNE	7 to 16
XXL	0301-04XXLE, 0301-04XXLEBN, 0301-04XXL20	10 to 22

Click'aV® and Click'aV Plus™ clips size	Compatible Click'aV® endo surgery clip removers	Compatible Click'aV® open surgery clip removers
M	0301-R804MLLE, 0301-R804MLLEB, 0301-R804MLLEI, 0301-R804MLLEIB, 0301-R804MLLEHS, 0301-R804MLLEHSB	0301-R804MMLL
ML	0301-R804MLLE, 0301-R804MLLEB, 0301-R804MLLEI, 0301-R804MLLEIB, 0301-R804MLLEHS, 0301-R804MLLEHSB	
L	0301-R804MLLE, 0301-R804MLLEB, 0301-R804MLLEI, 0301-R804MLLEIB, 0301-R804MLLEHS, 0301-R804MLLEHSB, 0301-R804LXLE, 0301-R804LXLEB, 0301-R804LXLEI, 0301-R804LXLEIB, 0301-R804LXLEHS, 0301-R804LXLEHSB – mostly recommended	0301-R804XL
XL	0301-R804LXLE, 0301-R804LXLEB, 0301-R804LXLEI, 0301-R804LXLEIB, 0301-R804LXLEHS, 0301-R804LXLEHSB	

Compatible with Grena Click'aV® and Click'aV Plus™ clips are also all appliers for polymer clips of other manufacturers with the same type and size of locking mechanism provided clip size corresponds with applier size. For best results it is highly recommended to use Grena appliers designed for Click'aV® and Click'aV Plus™ ligating clips.

- Warnings and precautions measures:**
- Any surgical and minimally invasive procedures should be performed only by persons having adequate training and familiarity with those techniques. Consult medical literature relative to techniques, complications, and hazards prior to performance of any surgical procedure.
 - Surgical instruments may vary from manufacturer to manufacturer. When surgical instruments and accessories from different manufacturers are employed together in a procedure, verify compatibility prior to initiation of the procedure. Failure to do this can result in inability to perform surgery.
 - Click'aV® and Click'aV Plus™ ligating clips are compatible with Click'aV® ligating clip appliers only and are not compatible with LigaV® or Vclip® clip appliers. Always ensure that correct Grena applier type was chosen prior to initiation of the procedure. Failure to do this can result in inability to perform surgery.
 - Grena recommends ligation of the renal artery in procedures other than laparoscopic live donor nephrectomy with more than one clip on the patient side with a minimum distal renal artery cuff of 2-3mm beyond the distal clip. If this is not followed, the clip may slip from the vessel during the procedure or in the postoperative period, leading to massive internal haemorrhage.
 - For the best results in sliding-clip technique Click'aV® ligating clips should be used. Click'aV Plus ligating clips are not recommended for this technique as they are designed for ultimate stability on the tissue, not the suture.
 - Surgeon is fully responsible to choose proper size of the clip and must determine how many clips is necessary to achieve satisfactory hemostasis and closure security.
 - Ligated structure sizes for each clip size is given for general information purposes only. Make sure clip's size is appropriate for the structure to be ligated. Clip should completely encompass the vessel or tissue structure. Tissue penetration by the clip latch may result in inability to close the clip, poor closure performance, slipping or opening of the clip during procedure or in postoperative period, leading to internal haemorrhage depending on the size of the vessel.
 - Do not use the clip or applier as a dissecting instrument. Applier tips may cause tissue injury or clip can fall out off the applier.
 - The clip must be latched to ensure proper ligation of the vessel or tissue. Inspect the ligation site after application to ensure proper closure of the clip. This should be repeated after the use of other surgical devices in the immediate area of the application.
 - Leave a distal cuff of tissue approximately 2-3 mm from the ligating clip if the tissue is to be divided, i.e. do not use the side of the clip as a cutting guide. If this is not followed, the clip may slip from the vessel during the procedure or in the postoperative period, leading to haemorrhage depending on the size of the vessel.
 - Do not squeeze the applier over other surgical instruments, staples, clips, gallstones or other hard structures as it may cause the clip to break.
 - Do not attempt to close the jaws on any tissue structure without a clip properly loaded into the jaws. Closure of empty jaws on a vessel or anatomic structure may result in patient injury.
 - Do not use damaged appliers. Using of damaged applier may result in improper location of a clip and cause the clip to fall off the tissue. When closed, jaw tips should be directly aligned and not offset. Always check the alignment of the applier jaws before use. If this is not done, patient injury may occur as the clip may be severely deformed during closure, preventing it from latching properly.
 - The following factors have serious influence on the closure of a clip: condition of an applier, force used by surgeon to close the clip, size of ligated tissue and features of the clip itself.
 - As for all other ligation technique it is required to check the place of ligation after applying a clip making sure it was located appropriately.

16. If endoscopic procedure is performed always confirm that the clip remains in the applier after insertion of the applier and clip through a cannula.
17. Always inspect the site for haemostasis before procedure is finished. Bleeding can be controlled by additional clips placement, electrocautery or surgical sutures.
18. Click'aV® and Click'aV Plus™ ligating clips can be opened with specially designed clip remover. It is highly recommended that remover be readily available during surgery involving the use of Click'aV® and Click'aV Plus™ ligating clips. Opened clip must be discarded and should not be reapplied again even if there are no any damages visible. Clip opened with the remover may develop microcracks and such clip could break or slip off the vessel leading to haemorrhage.
19. Grena does not promote or recommend any specific surgical practices. Surgical technique, types and sizes of tissues and vessels appropriate for ligation with Click'aV® and Click'aV Plus™ ligating clips are the responsibility of the surgeon.
20. Dispose of all opened clip cartridges no matter if all clips were used or not as sterility and full functionality of the device can be guaranteed if clips are used short after opening the package.
21. Implanted material is a non-absorbable Acetal Polymer. The material used does not require quantitative restrictions on the clips applied to the patient.
22. Use immediately after opening as storage of the devices post package opening leads to their contamination and creates a risk of an infection to the patient.
23. Take care to discard the product and packing after use, as well as unused but opened devices in accordance with hospital waste disposal practices and local regulations including, without limitation, those pertaining to human health and safety and the environment.
24. This product is intended for single patient and procedure use. Resterilization, reuse, reprocessing, modification may lead to serious consequences with death of patient included.
25. If any serious incident has occurred in relation to the device it should be reported to the manufacturer and the competent authority of the Member State, in which the user and/or patient is established.
26. Exercise caution when there is a potential for exposure to blood or bodily fluids. Adhere to hospital protocols regarding the use of protective wear and equipment.

	Keep dry	 eIFU indicator www.grena.co.uk/IFU	Consult electronic instructions for use		Manufacturer		Do not re-use
	Caution		Do not resterilize		Do not use if package is damaged and consult instructions for use		Use-by date
	Authorized Representative in the European Community		Catalogue number		Batch code		Quantity in package
	Sterilized using ethylene oxide		MR safe		Date of manufacture		Single sterile barrier system
	Medical device						

*The hard copies of instructions for use delivered with Grena products are always in english language.
If you require a hard copy of IFU in other language, you can contact Grena Ltd.
at ifu@grena.co.uk or + 44 115 9704 800.*

*Please scan the below QR code with the appropriate application.
It will connect you with Grena Ltd. website where you can choose eIFU in your preferable language.*

You can enter the website directly by typing in www.grena.co.uk/IFU in your browser.

*Make sure that paper version of IFU in your possession is in the latest revision prior to use of the device.
Always use the IFU in the latest revision.*

